



Gloss Meter

ETB-0833 Gloss Meter	
Brief Introduction	
The ETB-0833 GLOSSMETER is complete according with state standard GB9754-88、GB9966.5 and international standard ISO2813, all items of the performance is reaching to the state JJG696-2002 (Lens Luster Degree Instrument Measure Test Regulations) the first class work requirements.	
Technology parameter	
	ETB-0833
Projecting angle	20°60°85°
Applications	∴ The surface luster measure for printing ink, oil paint, bake lacquer, coat, woodwork and others. ∴ The surface luster measure for construction decoration materials: Marble; granite; glass chemical polishing brick, pottery brick ∴ The surface luster measure for plastic, sheet ∴ The surface luster measure for other non-metal materials
Measure range	0~200Gu
Measure range	< ±0.4Gu/30min
Value error	< ±1.2Gu
Power	DC1.5V×1
Environment temperature	0℃~40℃
Relative humidity	≤85%
Dimension	145×80×38mm