

0686 (60 °) Glos	smeter
Brief Introduction	
ISO2813, all items of	S METER is complete according with state standard GB9754-88、GB9966.5 and international standard the performance reach to the state JJG696-2002 (Lens Luster Degree Instrument Measure Test class work requirements.
Technology paramete	
	0686
Projecting angle	60 °
Applications	∴ The surface luster measure for printing ink, oil paint, bake lacquer, coat, woodwork and others. ∴ The surface luster measure for construction decoration materials: Marble; granite; glass chemical polishing brick, pottery brick and etc ∴ The surface luster measure for plastic, sheet and others ∴ The surface luster measure for other non-metal materials
Measure range	0~200Gu
Measure range	< ±0.4Gu/30min
Value error	< ±1.2Gu
Power	DC9V×1
Environment temperature	0°C~40°C
Relative humidity	≤85%
Dimension	114×70×38mm

